



Certificate of Analysis

Calibration against the 1st International standard for Activated Blood Coagulation Factor XI (FXIa), Human, (NIBSC code 13/100)

Factor XIa Control

| | | |
|---|----------------------------|---------|
| REF | Catalogue number | 1188 |
| LOT | Batch code | HHL2735 |
|  | Use by | 2019 09 |
|  | Upper limit of temperature | -20°C |

Activity after reconstitution with **4.0 mL** water:

Acceptance range
11.6 - 17.4 mIU/mL

The above Factor XIa activity apply only to **Lot#HHL2735**.

The calibration has been made against the 1st International standard for Activated Blood Coagulation Factor XI (FXIa), Human, NIBSC code 13/100 using the Rox Factor XIa kit Ref 110050.

The **Lot#HHL2735** has been approved for release after analyses according to established protocols with results demonstrating fulfillment of product specifications. All data concerning materials used, production data and release testing are documented and traceable under the individual Lot# which is stated on each package.



Pia Bryngelhed

Quality Control, Rossix AB

June 29, 2015

Date



Rossix AB
 Taljegardsgatan 3B
 SE-431 53 Molndal
 Sweden

Phone: +46 31 706 8965
 Fax: +46 31 706 8966
info@rossix.com
www.rossix.com