



Certificate of Analysis

Factor XIa Calibrator

REF	Catalogue number	1199
LOT	Batch code	CAJ2025
	Use by	2017-07
	Upper limit of temperature	-20°C


Assigned activity after reconstitution in **4.0 mL** water:

68 mU/mL

The above Factor XIa activity apply only to **Lot#CAJ2025**.

The calibration has been made against NIBSC reference reagent for Activated Blood Coagulation Factor XI (FXIa), Human, 11/236 using the Rox Factor XIa kit (Catalog number 11 00 50)

The **Lot#CAJ2025** has been approved for release after analyses according to established protocols with results demonstrating fulfillment of product specifications. All data concerning materials used, production data and release testing are documented and traceable under the individual Lot# which is stated on each package.


 Pia Bryngelhed
 Quality Control, Rossix AB

August 13, 2012
 Date



Rossix AB
 Taljegardsgatan 3B
 SE-431 53 Molndal
 Sweden

Phone: +46 31 706 8965
 Fax: +46 31 706 8966
info@rossix.com
www.rossix.com