



ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQRH00AZ

TECHNOZYM D-Dimer ELISA Kit 96 T

REF 2599006
LOT RH42B00.01
 31 - May - 2015

Inhalt Content	Reagenz Reagent	LOT	
1 x 12	D-Dimer Plate	RKS41B0	31-Jan-2016
1 x 0.3mL	D-Dimer Conjugate	RKV31B0	31-Mai-2015
1x 80mL	Wash buffer Concentrate	RNA41B0	29-Feb-2016
1 x 90mL	Incubation buffer	RNB41B0	31-Jan-2016
1 x 12mL	TMB Substrate	RKN33B0	30-Sep-2015
1 x 12mL	Stop solution	RKK33B0	31-Oct-2015
1 x 0,5mL	D-Dimer Calibrator 1	1KT31B0	31-Mai-2015
1 x 0.5mL	D-Dimer Calibrator 2	2KT31B0	31-Mai-2015
1 x 0.5mL	D-Dimer Calibrator 3	3KT31B0	31-Mai-2015
1 x 0.5mL	D-Dimer Calibrator 4	4KT31B0	31-Mai-2015
1 x 0.5mL	D-Dimer Calibrator 5	5KT42B0	30-Apr-2016
1 x 0.5mL	D-Dimer High Control	RKY31B0	31-Mai-2015
1 x 0.5mL	D-Dimer Low Control	RKZ31B0	31-Mai-2015

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result
OD	Calibrator 1	> 1.0
	Calibrator 5	< 0.1
Verwendbarkeit Usability	High Control (RKY31B0)	881.4 – 1469.0 ng/mL
	Low Control (RKZ31B0)	208.0 – 346.7 ng/mL

Durchführung von
Prepared by

Michael RIHA

29. Apr 2014

Datum / Date:

Freigabe von
Approved by

Fikret NASUFI

29. Apr 2014

Datum / Date: