



**ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS**

DQRH00AZ

TECHNOZYM D-Dimer ELISA Kit 96T

REF 2599006

LOT RH43B00.01

30-Apr-2016

Inhalt Content	Reagenz Reagent	LOT	
1 x	D-Dimer Plate	RKS42B0	30-Nov-2016
1 x 0.3mL	D-Dimer Conjugate	RKV31B1	30-Nov-2016
1x 80mL	Wash buffer Concentrate	RNA42B0	31-Aug-2016
1 x 90mL	Incubation buffer	RNB43B0	30-Sep-2016
1 x 12mL	TMB Substrate	RKN41B0	31-Aug-2016
1 x 12mL	Stop solution	RKK43B0	31-Dec-2016
1 x 0.5mL	D-Dimer Calibrator 1	1KT31B1	31-Dec-2016
1 x 0.5mL	D-Dimer Calibrator 2	2KT31B1	31-Dec-2016
1 x 0.5mL	D-Dimer Calibrator 3	3KT31B1	31-Dec-2016
1 x 0.5mL	D-Dimer Calibrator 4	4KT31B1	31-Dec-2016
1 x 0.5mL	D-Dimer Calibrator 5	5KT42B0	30-Apr-2016
1 x 0.5mL	D-Dimer High Control	RKY31B1	31-Dec-2016
1 x 0.5mL	D-Dimer Low Control	RKZ31B1	31-Dec-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.05	OK
	Calibrator 5	< 0.1	0.01	OK
Verwendbarkeit Usability	High Control (RKY31B1)	881.4 – 1469.0 ng/mL	1176.8	OK
	Low Control (RKZ31B1)	208.0 – 346.7 ng/mL	277.9	OK

Durchführung von
Prepared by

Michael RIHA

03-FEB-2015
Datum / Date:

Freigabe von
Approved by

Fikret NASUFI

03-FEB-2015
Datum / Date: