



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ9500AZ.01

NaCl Solution 0,9% 25mL

REF 4847127  
LOT 9541B00.01  
⌚ 30-Apr-2019

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
Volumen volume	≥25mL	25.2	OK
Verwendbarkeit Usability Chromogenes Substrat: F IIa Chromogenic substrate: F IIa	CCN LOT 1P44B00	72.0 – 120.0%	100.6 OK
	CCA LOT 3P24B00	30.0 – 50.0%	36.3 OK
Etiketten und Barcode Überprüfung Label and bar code Verification		OK	

Durchführung von  
Prepared by

Alessandro BENIGNI

06-Jun-2014  
Datum / Date:

Freigabe von  
Approved by

Lucas HOLOVICZ

06-Jun-2014  
Datum / Date: