

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ0T00AZ
---	---	----------

TECHNOTHROMBIN® TGA Kit 3x16 T.

REF 5006010
LOT 0T32B00.01
 28 – Feb - 2015

Inhalt Content	Reagenz Reagent	LOT	
1x3mL	TGA buffer	9P01B01	31-Dec-2015
3x1.5mL	TGA substrat	8F31B00	28-Feb-2015
1x1.5mL	TGA reagent RD	7G31B00	30-Jun-2015
1x0.5ml	TGA reagent RC Low	7D31B00	31-Mar-2015
1x0.5mL	TGA reagent RC High	7F31B00	31-Mar-2015
1x0.5mL	TGA Thrombin Calibrator	5T13B00	30-Jun-2015
1x1mL	TGA Control High	1B12B01	30-Apr-2015
1x1mL	TGA Control Low	2B34B00	30-Jun-2015

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung RC High	TGA Control High (nM Peak Thrombin) LOT 1B12B01	316.0 – 586.0	447.6	OK
Recovery of controls RC High	TGA Control Low (nM Peak Thrombin) LOT 2B34B00	132.4 – 245.8	178.2	OK
Kontrollwiederfindung RC Low	TGA Control High (nM Peak Thrombin) LOT 1B12B01	231.0 – 429.0	279.0	OK
Recovery of controls RC Low	TGA Control Low (nM Peak Thrombin) LOT 2B34B00	69.4 – 128.7	71.8	OK
Kontrollwiederfindung RD	TGA Control High (nM Peak Thrombin) LOT 1B12B01	166.0 – 308.0	255.0	OK
Recovery of controls RD	TGA Control Low (nM Peak Thrombin) LOT 2B34B00	202.8 – 376.7	259.8	OK