



**ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS**

DQ0T00AZ

TECHNOTHROMBIN® TGA Kit 3x16 T.


REF 5006010

LOT 0T52B00.01

30-Apr-2016

Inhalt Content	Reagenz Reagent	LOT	
1x3mL	TGA buffer	9P31B01	30-Apr-2017
3x1.5mL	TGA substrat	8F41B01	30-Sep-2017
1x1.5mL	TGA reagent RD	7G51B00	31-Jan-2017
1x0.5ml	TGA reagent RC Low	7D51B00	28-Feb-2017
1x0.5mL	TGA reagent RC High	7F51B00	30-Apr-2017
1x0.5mL	TGA Thrombin Calibrator	5T41B	30-Nov-2016
1x1mL	TGA Control High	1B41B00	30-Apr-2016
1x1mL	TGA Control Low	2B43B00	30-Jun-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung RC High	TGA Control High (nM Peak Thrombin) LOT 1B41B00)	276 – 512	505.1	OK
Recovery of controls RC High	TGA Control Low (nM Peak Thrombin) LOT 2B43B00	137 – 255	184.7	OK
Kontrollwiederfindung RC Low	TGA Control High (nM Peak Thrombin) LOT 1B41B00)	190 – 352	324.0	OK
Recovery of controls RC Low	TGA Control Low (nM Peak Thrombin) LOT 2B43B00	48 - 88	58.7	
Kontrollwiederfindung RD	TGA Control High (nM Peak Thrombin) LOT 1B41B00)	161 – 299	187.5	OK
Recovery of controls RD	TGA Control Low (nM Peak Thrombin) LOT 2B43B00	239 – 443	278.9	OK

	<p style="text-align: center;">ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</p>	<p style="text-align: center;">DQ0T00AZ</p>
---	---	---

Durchführung von
Prepared by

Fikret NASUFI
Fikret NASUFI

03-NOV-2015
Datum / Date:

Freigabe von
Approved by

Lavdim NASUFI
Lavdim NASUFI

03-NOV-2015
Datum / Date: