

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ7C00AZ
---	---	----------

Ceveron[®] TGA Reagenz B (RB) 5x1 mL

REF 5006202
LOT 7C42B.01
 31 - Jul - 2016

Inhalt Content	Reagenz Reagent	LOT	
5x1 mL	Ceveron [®] TGA RB Reagenz	7C42B	31-Jul-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1B41B00	120 – 222	168.8	OK
	TGA Control Low (nM Peak Thrombin) LOT 2B43B00	21 – 39	38.4	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by

Fikret NASU 

17-Jul-2016
Datum / Date:

Freigabe von
Approved by

Michael RIHA 

17-Jul-2016
Datum / Date: