



Certificate of Analysis

FIXa Calibrator

| | | |
|---|---------------------|----------------|
| REF | Catalog number | 9599 |
| LOT | Batch code | GSL1516 |
|  | Use by | 2019 04 |
|  | Storage temperature | -20°C or below |


Assigned activity after reconstitution in 2.0 mL water:

6.8 mIU/mL

The above Factor IXa activity apply only to **lot# GSL1516**. The calibration has been made against the 1st IS FIXa 97/562 using the Rox FIX-A kit (Art No 95 00 30)

The above Lot# has been approved for release after analyses according to established protocols with results demonstrating fulfilment of product specifications.

All data concerning materials used, production data and release testing are documented and traceable under the individual Lot# which is stated on each package.



Pia Bryngelhed
Quality Control, Rossix AB

June 18, 2014

Date