

**CERTIFICATE OF ANALYSIS**

# ***CHROMOGENIX***

PRODUCT:	Antithrombin 1 x 25 IU	PRODUCT # :	81079639
EXP. DATE:	2015-05	LOT # :	N0534571
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	654171

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
Antithrombin Activity	25	30	IU/vial	30	PASS
Assay Slope @ 37°C	-0.0024	-0.0034		-0.0034	PASS
Intercept @ 37°C	0.40	0.60		0.50	PASS
r <sup>2</sup>	0.990	--		0.999	PASS

KIT COMPONENTS :			
NAME	LOT NUMBER	EXP. DATE	VOLUME
Antithrombin, 25 IU	N0534571	2015-05	1 ml

PREPARED BY: *M. Oettinger*

DATE: 6/5/2013

APPROVAL SIGNATURE: *D. Tencza*

DATE: 6/5/2013