

CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT:	Antithrombin 1 x 25 IU	PRODUCT #:	81079639
EXP. DATE:	2017-06	LOT #:	N0656880
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	654171

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
Antithrombin Activity	25	30	IU/vial	29	PASS
Assay Slope @ 37°C	-0.0024	-0.0034		-0.0032	PASS
Intercept @ 37°C	0.40	0.60		0.45	PASS
r ²	0.990	--		0.997	PASS

KIT COMPONENTS :			
NAME	LOT NUMBER	EXP. DATE	VOLUME
Antithrombin, 25 IU	N0656880	2017-06	1 ml

PREPARED BY: VPATEL p.m. Patel DATE: 7/16/2015 7-16-15

APPROVAL SIGNATURE: DTENCZA J. Dencza DATE: 7/16/2015 7/16/15