

CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT: Antithrombin 10 x 10 IU	PRODUCT #: 82072039
EXP. DATE: 2016-11	LOT #: N0540346
VENDOR: INSTRUMENTATION LABORATORY	ASSAYED PER QCM: 654023

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION PASS/FAIL
	MIN	MAX			
ANTITHROMBIN, 10 IU (654023)					
ANTITHROMBIN ACTIVITY	9	11	IU/vial	10	PASS

KIT COMPONENTS :

NAME	LOT NUMBER	EXP. DATE	VOLUME
ANTITHROMBIN 10 IU	N0540346	2016-11	1 ml

All human source material used to manufacture this product have been tested according to FDA guidelines and found to be non-reactive for: HBsAg, antibodies to HIV1 and 2, and antibodies to HCV.

PREPARED BY: *M. Oettinger Maria Oettinger* DATE: 6/2/2014 ⁶⁻²⁻¹⁴ ~~6-3-14~~ ⁶⁻¹¹⁻¹⁴

APPROVAL SIGNATURE: *D. Tencza P. Manza* DATE: 6/2/2014 6/2/14

Form QCM 654023-2, Rev 01/09