

CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT: <u>Antithrombin 10 x 10 IU</u>	PRODUCT # : <u>82072039</u>
EXP. DATE: <u>2017-06</u>	LOT # : <u>N1243557</u>
VENDOR: <u>INSTRUMENTATION LABORATORY</u>	ASSAYED PER QCM: <u>654023</u>

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
ANTITHROMBIN, 10 IU (654023)					
ANTITHROMBIN ACTIVITY	9	11	IU/vial	10	PASS

KIT COMPONENTS :

NAME	LOT NUMBER	EXP.DATE	VOLUME
ANTITHROMBIN 10 IU	N1243557	2017-06	1 ml

All human source material used to manufacture this product have been tested according to FDA guidelines and found to be non-reactive for: HBsAg, antibodies to HIV1 and 2, and antibodies to HCV.

PREPARED BY: *NPATEL* DATE: 1/8/2015

APPROVAL SIGNATURE: *D TENOZA* DATE: 1/8/2015