

**CERTIFICATE OF ANALYSIS**

# ***CHROMOGENIX***

PRODUCT: FXa, 10 x 71 nkat	PRODUCT # : 82098539
EXP. DATE: 2016-07	LOT # : N0148451
VENDOR: INSTRUMENTATION LABORATORY	ASSAYED PER QCM: 650515

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
<b>FACTOR Xa, 71 nkat (650515)</b>					
ACTIVITY	61	81	nkat/vial	74	PASS
STANDARD CURVE (r <sup>2</sup> )	0.990	1.000	--	1.000	PASS
CONTROL RECOVERY	66.1	73.1	nkat/vial	66.8	PASS

**KIT COMPONENTS :**

NAME	LOT NUMBER	EXP. DATE	VOLUME
FXa, 71 nkat	N0148451	2016-07	2 ml

PREPARED BY: *N. Patel* DATE: 2/27/2014

APPROVAL SIGNATURE: *D. Tencza* DATE: 2/27/2014