

CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT: FXa, 10 x 71 nkat	PRODUCT #: 82098539
EXP. DATE: 2017-08	LOT #: N0254681
VENDOR: INSTRUMENTATION LABORATORY	ASSAYED PER QCM: 650515

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION PASS/FAIL
	MIN	MAX			
FACTOR Xa, 71 nkat (850515)					
ACTIVITY	61	81	nkat/vial	74	PASS
STANDARD CURVE (r^2)	0.990	1.000	-	1.000	PASS
CONTROL RECOVERY	67.5	74.6	nkat/vial	70.1	PASS

KIT COMPONENTS :

NAME	LOT NUMBER	EXP. DATE	VOLUME
FXa, 71 nkat	N0254681	2017-08	2 ml

PREPARED BY: Maria Cettinzer DATE: 2-25-15

APPROVAL SIGNATURE: P. Anza DATE: 2/25/15