

CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT: <u>FXa, 10 x 71 nkat</u>	PRODUCT # : <u>82098539</u>
EXP. DATE: <u>2015-10</u>	LOT # : <u>N0433802</u>
VENDOR: <u>INSTRUMENTATION LABORATORY</u>	ASSAYED PER QCM: <u>650515</u>

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
FACTOR Xa, 71 nkat (650515)					
ACTIVITY	61	81	nkat/vial	68	PASS
STANDARD CURVE (r ²)	0.990	1.000	--	1.000	PASS
CONTROL RECOVERY	64.9	71.7	nkat/vial	67.3	PASS

KIT COMPONENTS :

NAME	LOT NUMBER	EXP. DATE	VOLUME
FXa, 71 nkat	N0433802	2015-10	2 ml

PREPARED BY: *VPATEL*

DATE: 5/6/2013

APPROVAL SIGNATURE: *D7ENCZA*

DATE: 5/6/2013