

CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT: FXa, 10 x 71 nkat	PRODUCT #: 82098539
EXP. DATE: 2016-11	LOT #: N0540424
VENDOR: INSTRUMENTATION LABORATORY	ASSAYED PER QCM: 650515

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
FACTOR Xa, 71 nkat (650515)					
ACTIVITY	61	81	nkat/vial	69	PASS
STANDARD CURVE (r^2)	0.990	1.000	--	0.998	PASS
CONTROL RECOVERY	66.1	73.1	nkat/vial	69.0	PASS

KIT COMPONENTS :

NAME	LOT NUMBER	EXP. DATE	VOLUME
FXa, 71 nkat	N0540424	2016-11	2 ml

PREPARED BY: M. Oettinger Marie Oettinger DATE: 6/12/2014 6/12/14

APPROVAL SIGNATURE: D. Tencza D. Tencza DATE: 6/12/2014 6/12/14