

CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT: FXa, 10 x 71 nkat	PRODUCT # : 82098539
EXP. DATE: 2017-06	LOT # : N1243555
VENDOR: INSTRUMENTATION LABORATORY	ASSAYED PER QCM: 650515

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
FACTOR Xa, 71 nkat (650515)					
ACTIVITY	61	81	nkat/vial	68	PASS
STANDARD CURVE (r ²)	0.990	1.000	--	1.000	PASS
CONTROL RECOVERY	66.1	73.1	nkat/vial	68.3	PASS

KIT COMPONENTS :

NAME	LOT NUMBER	EXP. DATE	VOLUME
FXa, 71 nkat	N1243555	2017-06	2 ml

PREPARED BY: VPATEL DATE: 1/9/2015

APPROVAL SIGNATURE: DHARJ DATE: 1/9/2015