

CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT:	COATEST APC Resistance Control Plasma Level 1	PRODUCT # :	82265063
EXP. DATE:	2015-12	LOT # :	N1237730
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	C00423

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION	
	MIN	MAX				
Analysis with COATEST APC R:						
APT Time	CaCl ₂	28	34	seconds	30	PASS
	APC/CaCl ₂	90	--	seconds	123	PASS
APC Ratio		3.5	4.4	--	4.1	PASS

Analysis with COATEST APC R V:						
APT Time	CaCl ₂	--	38	seconds	34	PASS
	APC/CaCl ₂	75	--	seconds	109	PASS
APC Ratio		2.7	3.2	--	3.2	PASS

Kit Components:

DESCRIPTION	LOT NUMBER	EXP. DATE	VOLUME
Control Plasma Level 1	N1237730	2015-12	1 ml (x 5)

We hereby certify that the above product(s):

- have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;
- meet all applicable requirements of the 98/79/EC Directive according to which they are CE marked.

PREPARED BY:

N. Patel

N. M. Patel

DATE:

12/16/2013

APPROVAL SIGNATURE:

D. Tencza

D. Tencza

DATE:

12/16/13
12/16/2013