

CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT:	COATEST APC Resistance Control Plasma Level 1	PRODUCT #:	82265063
EXP. DATE:	2016-12	LOT #:	N1243568
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	C00423

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION	
Analysis with COATEST APC R:						
APT Time	CaCl ₂	MIN 28	MAX 34	seconds	30	PASS
	APC/CaCl ₂	90	-	seconds	119	PASS
APC Ratio		3.5	4.4	-	4.0	PASS

Analysis with COATEST APC R V:						
APT Time	CaCl ₂	-	38	seconds	36	PASS
	APC/CaCl ₂	75	-	seconds	110	PASS
APC Ratio		2.7	3.2	-	3.1	PASS

Kit Components:

DESCRIPTION	LOT NUMBER	EXP. DATE	VOLUME
Control Plasma Level 1	N1243568	2016-12	1 ml (x 5)

We hereby certify that the above product(s):

- have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;
- meet all applicable requirements of the 98/79/EC Directive according to which they are CE marked.

PREPARED BY: NPATEL N m. Patel

DATE: 1/15/2015 1:15:15

APPROVAL SIGNATURE: DTENCZA D. Tenca

DATE: 1/15/2015 1/15/15