CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT:	COATEST APC Resistance Control Plasma Level 2	PRODUCT # :	82266863
EXP. DATE:	2017-01	LOT # :	N0154142
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	H01444

TEST DESCRIPTION		SPECIFICATIONS			RESULTS	DISPOSITION
Analysis with COATEST APC R:		**MIN**	**MAX**	**UNITS**		**PASS/FAIL**
APT Time	$CaCl_2$	27	44	seconds	40	PASS
	APC/$CaCl_2$	--	101	seconds	90	PASS
APC Ratio		--	2.3	--	2.2	PASS
Analysis with COATEST APC R V:						
APT Time	$CaCl_2$	27	40	seconds	34	PASS
APC Ratio		1.4	1.9	--	1.8	PASS

Kit Components:

DESCRIPTION	LOT NUMBER	EXP.DATE	VOLUME
Control Plasma Level 2	N0154142	2017-01	1 ml (x 5)

We hereby certify that the above product(s):
· have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;
· meet all applicable requirements of the 98/79/EC Directive according to which they are CE marked.

PREPARED BY: N. Patel N M. Patel DATE: 2/17/2015 2·19·15

APPROVAL SIGNATURE: D. Teneza D. Teneza DATE: 2/17/15 2/17/2015

526 Route 303, Orangeburg, New York 10962, Telephone 845-365-8000