

CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT:	COATEST APC Resistance Control Plasma Level 2	PRODUCT # :	82266863
EXP. DATE:	2016-05	LOT # :	N0540174
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	H01444

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION	
	MIN	MAX				
Analysis with COATEST APC R:						
APT Time					PASS/FAIL	
	<u>CaCl₂</u>	27	44	seconds	40	PASS
	<u>APC/CaCl₂</u>	--	101	seconds	83	PASS
APC Ratio		--	2.3	--	2.1	PASS

Analysis with COATEST APC R V:						
APT Time	<u>CaCl₂</u>	27	40	seconds	33	PASS
APC Ratio		1.4	1.9	--	1.7	PASS

Kit Components:

DESCRIPTION	LOT NUMBER	EXP. DATE	VOLUME
Control Plasma Level 2	N0540174	2016-05	1 ml (x 5)

We hereby certify that the above product(s):
 · have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;
 · meet all applicable requirements of the 98/79/EC Directive according to which they are CE marked.

PREPARED BY: MPATEL N.m. Patel DATE: 5-21-14
 APPROVAL SIGNATURE: DTENCZA J. Dencza DATE: 5/21/14