

CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT:	COATEST APC Resistance Control Plasma Level 2	PRODUCT # :	82266863
EXP. DATE:	2015-06	LOT # :	N0635141
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	H01444

TEST DESCRIPTION	SPECIFICATIONS		RESULTS	DISPOSITION	
Analysis with COATEST APC R:	MIN	MAX	UNITS	PASS/FAIL	
APT Time <u>CaCl₂</u>	27	44	seconds	39	PASS
<u>APC/CaCl₂</u>	--	101	seconds	80	PASS
APC Ratio	--	2.3	--	2.0	PASS

Analysis with COATEST APC R V:	MIN	MAX	UNITS	RESULTS	DISPOSITION
APT Time <u>CaCl₂</u>	27	40	seconds	36	PASS
APC Ratio	1.4	1.9	--	1.6	PASS

Kit Components:

DESCRIPTION	LOT NUMBER	EXP. DATE	VOLUME
Control Plasma Level 2	N0635141	2015-06	1 ml (x 5)

We hereby certify that the above product(s):
 - have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;
 - meet all applicable requirements of the 98/79/EC Directive according to which they are CE marked.

PREPARED BY: N. Patel *N. Patel* DATE: 8/15/2013 *8-15-13*

APPROVAL SIGNATURE: D. Tencza *J. Longza* DATE: 8/15/2013 *8/20/13*