

**CERTIFICATE OF ANALYSIS**

**CHROMOGENIX**

PRODUCT:	COATEST APC Resistance Control Plasma Level 2	PRODUCT # :	82266863
EXP. DATE:	2017-09	LOT # :	N0958710
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	H01444

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	DISPOSITION
	MIN	MAX	UNITS		
<b>Analysis with COATEST APC R:</b>					<b>PASS/FAIL</b>
APT Time <u>CaCl<sub>2</sub></u>	27	44	seconds	40	PASS
<u>APC/CaCl<sub>2</sub></u>	--	101	seconds	74	PASS
APC Ratio	--	2.3	--	1.9	PASS

<b>Analysis with COATEST APC R V:</b>					
APT Time <u>CaCl<sub>2</sub></u>	27	40	seconds	39	PASS
APC Ratio	1.4	1.9	--	1.7	PASS

**Kit Components:**

DESCRIPTION	LOT NUMBER	EXP. DATE	VOLUME
Control Plasma Level 2	N0958710	2017-09	1 ml (x 5)

We hereby certify that the above product(s):

- have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;
- meet all applicable requirements of the 98/79/EC Directive according to which they are CE marked.

PREPARED BY: WPA TEL N. m. Patel

DATE: 10/16/2015 10:16:15

APPROVAL SIGNATURE: DTENCZA P. Danga

DATE: 10/16/2015 10/16/15