CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT:	Control Plasma-LMW Heparin	PRODUCT # :	823492
EXP. DATE:	2015-09	LOT # :	N0333372
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	823492/823500

TEST DESCRIPTION		SPECIFICATIONS			RESULTS	DISPOSITION
A. Control Plasma, Low: LMW Heparin (H01337)		**MIN**	**MAX**	**UNITS**		**PASS/FAIL**
Mean value		0.33	0.42	IU/vial	0.38	PASS
Intra series CV (n=60)		-	10	%	7	PASS
B. Control Plasma, High, LMW Heparin (H01346)						
Mean value		0.64	0.84	IU/vial	0.72	PASS
Intra series CV (n=60)		-	5	%	4	PASS

Kit Components:

Name	Lot Number	Exp. Date	Mfg. Date
Control Plasma, Low, LMW Heparin	N0333367	2015-09	3/12/2013
Control Plasma, High, LMW Heparin	N0333368	2015-09	3/12/2013

THE PLASMA MATERIAL USED IN THE MANUFACTURE OF THIS PRODUCT HAS BEEN TESTED AND FOUND NON-REACTIVE FOR HIV, HBsAg AND ANTI-HCV.

We hereby certify that the above product(s):
- have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;
- meet all applicable requirements of the 98/79/EC Directive according to which they are CE marked.

PREPARED BY: N. Patel N.M. Patel DATE: 3/21/2013 3.21.13

APPROVAL SIGNATURE: D. Tenza D. Tenza DATE: 3/21/2013 3/21/13

Form QCM0082350063/0082349263-1, Rev. 09/10