



Certificate of Analysis

REAADS
Protein C Antigen

Kit Lot: PC-152
Plate Lot #: PCW-055
Product #: 035-001
Expiration Date: 2015-08-25

Quality Control Results

Protein C Antigen		Testing Date: 2014-03-07	
Attribute	Value	Specification	Result by Assay
Reference Plasma 100% OD	101 % of Normal	≥ 1.0	1.229
Reagent Background O.D.		≤ 0.100	0.018
Reference Curve R-Squared		≥ 0.95	0.98033
Control Recovery		Label Range	Result by Assay
Normal Control		73 - 141 %	103.8 %
Deficient Control		28 - 44 %	40.0 %

This signed Certificate of Analysis affirms that the product has met the requirements of the relevant specifications as set by Corgenix, Inc.

Human sera used to prepare the controls included in this kit have been tested and shown to be non-reactive for HBsAg and negative for antibodies to HCV and HIV 1 and 2 by FDA required tests. However, all human blood derivatives, including patient samples, should be handled as though capable of transmitting disease.

This product is warranted to perform as described in the package insert. Corgenix, Inc. disclaims any implied warranty of merchantability or fitness for a particular use, and in no event shall Corgenix, Inc. be liable for consequential damage.

Quality Assurance Approval By: Gladys Reese

Signed: 2014-05-07
Title: QA Associate

Corgenix, Inc.
11575 Main Street, Suite 400
Broomfield, CO 80020
Phone: (303) 457-4345
Fax: (303) 457-4519