



Certificate of Analysis

REAADS
VWF Activity ELISA

Kit Lot: VA-087
Plate Lot #: VAW-032
Product #: 10826
Expiration Date: 2016-10-07

Quality Control Results


VWF Activity		Testing Date: 2015-05-04		
Attribute	Value	Specification	Result by Assay	
Reference Plasma 100% OD	82 % of Normal	≥ 1.0	1.249	
Reagent Background O.D.		≤ 0.100	0.009	
Reference Curve R-Squared		≥ 0.95	0.98890	
Control Recovery		Label Range	Result by Assay	
Normal Control	78 - 116 %		104.0 %	
Deficient Control	25 - 38 %		35.2 %	

This signed Certificate of Analysis affirms that the product has met the requirements of the relevant specifications as set by Corgenix, Inc.

Human sera used to prepare the controls included in this kit have been tested and shown to be nonreactive for HBsAG and negative for antibodies to HCV and HIV 1 and 2 by FDA required tests. However, all human blood derivatives, including patient samples, should be handled as though capable of transmitting disease.

This product is warranted to perform as described in the package insert. Corgenix, Inc. disclaims any implied warranty of merchantability or fitness for a particular use, and in no event shall Corgenix, Inc. be liable for consequential damage.

Quality Assurance Approval By: **Gladys Reese**

Signed:  2015-05-11
Title: **QA Associate**

Corgenix, Inc.
11575 Main Street, Suite 400
Broomfield, CO 80020
Phone: (303) 457-4345
Fax: (303) 457-4519