

CERTIFICATE OF ANALYSIS

**CHROMOGENIX**

PRODUCT:	COAMATIC FACTOR VIII	PRODUCT #:	82258563
EXP. DATE:	2016-06	LOT #:	N0456134
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	822585/49730503

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	PASS/FAIL
	MIN	MAX			
<b>HIGH RANGE</b>					
Slope	0.33	0.51	--	0.44	PASS
Intercept	--	0.12	--	0.06	PASS
r <sup>2</sup>	0.990	--	--	0.997	PASS
<b>LOW RANGE</b>					
Slope	2.8	5.8	--	5.5	PASS
Intercept	--	0.18	--	0.16	PASS
r <sup>2</sup>	0.990	--	--	0.998	PASS

Kit Components:

Name	Lot Number	Exp. date
S-2765/ I-2581	N0254972	2017-02
Factor Reagent	N0640720	2016-06
Buffer	N0641808	2016-08

We hereby certify that the above product(s):

- have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;
- meet all applicable requirements of the 98/79/EC Directive according to which they are CE marked.

PREPARED BY: NPATEL N.m. Patel DATE: 4.30.15  
 APPROVAL SIGNATURE: D TENCZA J. Tencza DATE: 4/30/15