

CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT:	<u>COAMATIC FACTOR VIII</u>	PRODUCT # :	<u>82258563</u>
EXP. DATE:	<u>2015-09</u>	LOT # :	<u>N1036805</u>
VENDOR:	<u>INSTRUMENTATION LABORATORY</u>	ASSAYED PER QCM:	<u>822585/49730503</u>

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	PASS/FAIL
	MIN	MAX			
HIGH RANGE					
Slope	0.33	0.51	--	0.42	PASS
Intercept	--	0.12	--	0.04	PASS
r ²	0.990	--	--	0.994	PASS
LOW RANGE					
Slope	2.8	5.8	--	4.5	PASS
Intercept	--	0.18	--	0.10	PASS
r ²	0.990	--	--	0.996	PASS

Kit Components:

Name	Lot Number	Exp. date
S-2765/ I-2581	N1036713	2015-10
Factor Reagent	N0936602	2015-09
Buffer	N0936393	2015-09

We hereby certify that the above product(s):

- have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;
- meet all applicable requirements of the 98/79/EC Directive according to which they are CE marked.

PREPARED BY: NPATEL DATE: 10/22/2013

APPROVAL SIGNATURE: DTENCZA DATE: 10/22/2013