CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT:	COATEST APC Resistance Kit	PRODUCT # :	82264363
EXP. DATE:	2016-01	LOT # :	N1042840
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	822643

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	
	MIN	MAX	UNITS		Pass/Fail
APC-ratio on ACL					
Control Plasma Level 1 (Lot:__ N1238033	3.5	4.4	--	4.1	PASS
Control Plasma Level 2 (Lot:__ N0132841	--	2.3	--	1.8	PASS

Kit Components:

Name	Lot Number	Exp. date	Volume
$CaCl_2$	N0741332	2017-07	8 ml (x 1)
APTT Reagent	N0741426	2016-01	16 ml (x 1)
APC/$CaCl_2$	N0942504	2016-03	2 ml (x 4)

We hereby certify that the above product(s):
- have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;
- meet all applicable requirements of the 98/79/EC Directive according to which they are CE marked.

PREPARED BY: N. Patel DATE: 10/17/2014

APPROVAL SIGNATURE: D. Tencza DATE: 10/17/2014

Form QCM0082264363-1, Rev. 05/07

526 Route 303, Orangeburg, New York 10962, Telephone 845-365-8000