



Certificate of Analysis

PRODUCT:	CoaTest SP Factor FVIII Kit	PRODUCT # :	82408663
EXP. DATE:	2017-09	LOT # :	N1058869
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	824086

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
Coatest SP Factor VIII Kit (82103363)					
High Range					
Slope	0.0077	0.0112	--	0.0107	PASS
Intercept	--	0.250	--	0.133	PASS
Standard Curve (r)	0.99	--	--	0.999	PASS
Low Range					
Slope	0.020	0.060	--	0.043	PASS
Intercept	--	0.406	--	0.130	PASS
Standard Curve (r)	0.99	--	--	0.997	PASS

KIT COMPONENTS :			
NAME	LOT NUMBER	EXP.DATE	VOLUME
CaCl ₂	N0958584	2018-09	6 ml
CT SP Substrate	N0958434	2017-09	12 ml
Factor Reagent	N1058868	2018-04	10 ml
CT SP Buffer, pH 7.3	N0958711	2017-09	20 ml
CT SP Phospholipid	N0958714	2017-09	2 ml

PREPARED BY: MPATEL

DATE: 10/30/15

APPROVAL SIGNATURE: DTENCZA

DATE: 10/30/15