



Certificate of Analysis

PRODUCT: CoaTest SP Factor FVIII Kit	PRODUCT # : 82408663
EXP. DATE: 2015-10	LOT # : N1137694
VENDOR: INSTRUMENTATION LABORATORY	ASSAYED PER QCM: 824086

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
Coatest SP Factor VIII Kit (82103363)					
High Range					
Slope	0.0077	0.0112	--	0.0091	PASS
Intercept	--	0.250	--	0.139	PASS
Standard Curve (r)	0.99	--	--	0.998	PASS
Low Range					
Slope	0.020	0.060	--	0.041	PASS
Intercept	--	0.406	--	0.209	PASS
Standard Curve (r)	0.99	--	--	0.998	PASS

KIT COMPONENTS :			
NAME	LOT NUMBER	EXP. DATE	VOLUME
CaCl ₂	N1037020	2016-10	6 ml
CT SP Substrate	N1037140	2015-10	12 ml
Factor Reagent	N0534294	2015-11	10 ml
CT SP Buffer, pH 7.3	N1037120	2015-10	20 ml
CT SP Phospholipid	N1037141	2015-10	2 ml

PREPARED BY: NPATEL

DATE: 12/19/13

APPROVAL SIGNATURE: DTENCZA

DATE: 12/19/13