



## Certificate of Analysis

PRODUCT:	CoaTest SP4 FVIII Kit	PRODUCT #:	82409463
EXP. DATE:	2015-10	LOT #:	NO248842
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER OCM:	824094

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
Coatest SP4 Factor VIII Kit (82409463)					
High Range					
Slope	0.0077	0.0112	--	0.0083	PASS
Intercept	--	0.250	--	0.110	PASS
Standard Curve (r)	0.99	--	--	0.994	PASS
Low Range					
Slope	0.020	0.060	--	0.036	PASS
Intercept	--	0.406	--	0.185	PASS
Standard Curve (r)	0.99	--	--	0.992	PASS

KIT COMPONENTS :			
NAME	LOT NUMBER	EXP. DATE	VOLUME
CaCl <sub>2</sub>	N1037020	2016-10	(1x) 6 ml
CT SP Substrate	N1237699	2015-12	(1x) 12 ml
Factor Reagent	N1137689	2016-05	(4x) 3 ml
CT SP Buffer, pH 7.3	N1037120	2015-10	(1x) 20 ml
CT SP Phospholipid	N1037141	2015-10	(1x) 2 ml

PREPARED BY: *P. M. Patel* *N. Patel* DATE: *3-31-14* **03/31/14**

APPROVAL SIGNATURE: *D. Tenaga* *D. Inga* DATE: *3/31/14* **03/31/14**

Form QCM0082409463-1, Rev. 02/10