

CHROMOGENIX

Certificate of Analysis

PRODUCT: CoaTest SP4 FVIII Kit	PRODUCT #: 82409463
EXP. DATE: 2015-03	LOT #: N0333454
VENDOR: INSTRUMENTATION LABORATORY	ASSAYED PER QCM: 824094

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
Coatest SP4 Factor VIII Kit (82409463)					
High Range					
Slope	0.0077	0.0112	--	0.0092	PASS
Intercept	--	0.250	--	0.180	PASS
Standard Curve (r)	0.99	--	--	0.999	PASS
Low Range					
Slope	0.020	0.060	--	0.040	PASS
Intercept	--	0.406	--	0.367	PASS
Standard Curve (r)	0.99	--	--	0.997	PASS

KIT COMPONENTS :

NAME	LOT NUMBER	EXP. DATE	VOLUME
CaCl ₂	N0233032	2016-02	(1x) 6 ml
CT SP Substrate	N0333453	2015-03	(1x) 12 ml
Factor Reagent	N0433806	2015-10	(4x) 3 ml
CT SP Buffer, pH 7.3	N0333377	2015-03	(1x) 20 ml
CT SP Phospholipid	N0333365	2015-03	(1x) 2 ml

PREPARED BY: N.M. Patel DATE: 4-29-13

APPROVAL SIGNATURE: DTENCZA V. Danga DATE: 04/29/13 4/29/13