

# CHROMOGENIX

## Certificate of Analysis

PRODUCT: CoaTest SP4 FVIII Kit	PRODUCT #: 82409463
EXP. DATE: 2016-08	LOT #: N0355378
VENDOR: INSTRUMENTATION LABORATORY	ASSAYED PER QCM: 824094

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
<b>Coatest SP4 Factor VIII Kit (82409463)</b>					
<b>High Range</b>					
Slope	0.0077	0.0112	--	0.0098	PASS
Intercept	--	0.250	--	0.158	PASS
Standard Curve (r)	0.99	--	--	0.995	PASS
<b>Low Range</b>					
Slope	0.020	0.060	--	0.046	PASS
Intercept	--	0.406	--	0.235	PASS
Standard Curve (r)	0.99	--	--	0.999	PASS

KIT COMPONENTS :			
NAME	LOT NUMBER	EXP. DATE	VOLUME
CaCl <sub>2</sub>	N0841692	2017-08	(1x) 6 ml
CT SP Substrate	N0942242	2016-09	(1x) 12 ml
Factor Reagent	N0254967	2017-08	(4x) 3 ml
CT SP Buffer, pH 7.3	N0841953	2016-08	(1x) 20 ml
CT SP Phospholipid	N0842034	2016-08	(1x) 2 ml

PREPARED BY: TRATEL N.M. Patel DATE: 04/08/15 4:8:15

APPROVAL SIGNATURE: DTENCZA J. Dencza DATE: 04/08/15 4/9/15  
DNT  
4/10/15

Form QCM0082409463-1, Rev. 02/10