

# CHROMOGENIX

## Certificate of Analysis

PRODUCT: CoaTest SP4 FVIII Kit	PRODUCT #: 82409463
EXP. DATE: 2015-04	LOT #: N0434108
VENDOR: INSTRUMENTATION LABORATORY	ASSAYED PER QCM: 824094

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	DISPOSITION
	MIN	MAX	UNITS		
<b>Coatest SP4 Factor VIII Kit (82409463)</b>					
<b>High Range</b>					
Slope	0.0077	0.0112	--	0.0093	PASS
Intercept	--	0.250	--	0.120	PASS
Standard Curve (r)	0.99	--	--	0.999	PASS
<b>Low Range</b>					
Slope	0.020	0.060	--	0.040	PASS
Intercept	--	0.406	--	0.198	PASS
Standard Curve (r)	0.99	--	--	0.999	PASS

### KIT COMPONENTS :

NAME	LOT NUMBER	EXP. DATE	VOLUME
CaCl <sub>2</sub>	N0434006	2016-04	(1x) 6 ml
CT SP Substrate	N0434107	2015-04	(1x) 12 ml
Factor Reagent	N0434010	2015-10	(4x) 3 ml
CT SP Buffer, pH 7.3	N0434008	2015-04	(1x) 20 ml
CT SP Phospholipid	N0434009	2015-04	(1x) 2 ml

PREPARED BY: *N. M. Patel* 5-13-13  
*NPATEL* DATE: 05/13/13  
 APPROVAL SIGNATURE: *DTENCZA* 5-14-13  
*DTENCZA* DATE: 05/13/13