



Certificate of Analysis

PRODUCT:	CoaTest SP4 FVIII Kit	PRODUCT #:	82409463
EXP. DATE:	2015-04	LOT #:	N0735496
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	824094

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
Coatest SP4 Factor VIII Kit (82409463)					
High Range					
Slope	0.0077	0.0112	--	0.0083	PASS
Intercept	--	0.250	--	0.109	PASS
Standard Curve (r)	0.99	--	--	0.999	PASS
Low Range					
Slope	0.020	0.060	--	0.040	PASS
Intercept	--	0.406	--	0.208	PASS
Standard Curve (r)	0.99	--	--	1.000	PASS

KIT COMPONENTS :

NAME	LOT NUMBER	EXP. DATE	VOLUME
CaCl ₂	N0434006	2016-04	(1x) 6 ml
CT SP Substrate	N0434107	2015-04	(1x) 12 ml
Factor Reagent	N0735494	2016-01	(4x) 3 ml
CT SP Buffer, pH 7.3	N0434008	2015-04	(1x) 20 ml
CT SP Phospholipid	N0434009	2015-04	(1x) 2 ml

PREPARED BY: N. M. Patel DATE: 8-22-13
NPATEL DATE: 08/22/13

APPROVAL SIGNATURE: DTENCZA DATE: 08/22/13 8/23/13
D. Tenca