



Certificate of Analysis

PRODUCT: CoaTest SP4 FVIII Kit	PRODUCT #: 82409463
EXP. DATE: 2017-09	LOT #: N0958715
VENDOR: INSTRUMENTATION LABORATORY	ASSAYED PER QCM: 824094

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
Coatest SP4 Factor VIII Kit (82409463)					
High Range					
Slope	0.0077	0.0112	--	0.0107	PASS
Intercept	--	0.250	--	0.118	PASS
Standard Curve (r)	0.99	--	--	0.998	PASS
Low Range					
Slope	0.020	0.060	--	0.047	PASS
Intercept	--	0.406	--	0.124	PASS
Standard Curve (r)	0.99	--	--	0.999	PASS

KIT COMPONENTS :			
NAME	LOT NUMBER	EXP. DATE	VOLUME
CaCl ₂	N0958584	2018-09	(1x) 6 ml
CT SP Substrate	N0958434	2017-09	(1x) 12 ml
Factor Reagent	N0958331	2018-03	(4x) 3 ml
CT SP Buffer, pH 7.3	N0958711	2017-09	(1x) 20 ml
CT SP Phospholipid	N0958714	2017-09	(1x) 2 ml

PREPARED BY:

MPATEL N.M. Patel

DATE:

10/29/15 10/29/15

APPROVAL SIGNATURE:

DTENCZA D. Dencza

DATE:

10/29/15 10/29/15