

CHROMOGENIX

Certificate of Analysis

PRODUCT:	CoaTest SP4 FVIII Kit	PRODUCT # :	82409463
EXP. DATE:	2015-10	LOT # :	N1137476
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	824094

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
Coatest SP4 Factor VIII Kit (82409463)					PASS/FAIL
High Range					
Slope	0.0077	0.0112	--	0.0092	PASS
Intercept	--	0.250	--	0.123	PASS
Standard Curve (r)	0.99	--	--	0.998	PASS
Low Range					
Slope	0.020	0.060	--	0.042	PASS
Intercept	--	0.406	--	0.202	PASS
Standard Curve (r)	0.99	--	--	0.998	PASS

KIT COMPONENTS :

NAME	LOT NUMBER	EXP. DATE	VOLUME
CaCl ₂	N1037020	2016-10	(1x) 6 ml
CT SP Substrate	N1037140	2015-10	(1x) 12 ml
Factor Reagent	N1036828	2016-04	(4x) 3 ml
CT SP Buffer, pH 7.3	N1037120	2015-10	(1x) 20 ml
CT SP Phospholipid	N1037141	2015-10	(1x) 2 ml

PREPARED BY: N. M. Patel DATE: 11-26-13
NPATEL
 APPROVAL SIGNATURE: D. Tenca DATE: 11/26/13
D. Tenca