CHROMOGENIX

Certificate of Analysis

PRODUCT:	CoaTest SP4 FVIII Kit	PRODUCT # :	82409463
EXP. DATE:	2015-10	LOT # :	N1237797
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	824094

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	DISPOSITION
	MIN	MAX	UNITS		PASS/FAIL
Coatest SP4 Factor VIII Kit (82409463)					
High Range					
Slope	0.0077	0.0112	--	0.0088	PASS
Intercept	--	0.250	--	0.122	PASS
Standard Curve (r)	0.99	--	--	0.999	PASS
Low Range					
Slope	0.020	0.060	--	0.037	PASS
Intercept	--	0.406	--	0.198	PASS
Standard Curve (r)	0.99	--	--	0.999	PASS

KIT COMPONENTS :

NAME	LOT NUMBER	EXP.DATE	VOLUME
$CaCl_2$	N1037020	2016-10	(1x) 6 ml
CT SP Substrate	N1237699	2015-12	(1x) 12 ml
Factor Reagent	N1137242	2016-05	(4x) 3 ml
CT SP Buffer, pH 7.3	N1037120	2015-10	(1x) 20 ml
CT SP Phospholipid	N1037141	2015-10	(1x) 2 ml

N.M. Patel 3-31-14

PREPARED BY: NPATEL DATE: 03/27/14

3/31/14

APPROVAL SIGNATURE: DTENCZA DTenca DATE: 03/27/14

Form QCM0082409463-1, Rev. 02/10

526 Route 303, Orangeburg, New York 10962, Telephone 845-365-8000