



Certificate of Analysis

PRODUCT:	CoaTest SP4 FVIII Kit	PRODUCT #:	82409463
EXP. DATE:	2017-03	LOT #:	N0456087
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	824094

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
Coatest SP4 Factor VIII Kit (82409463)					
High Range					
Slope	0.0077	0.0112	--	0.0094	PASS
Intercept	--	0.250	--	0.127	PASS
Standard Curve (r)	0.99	--	--	0.997	PASS
Low Range					
Slope	0.020	0.060	--	0.040	PASS
Intercept	--	0.406	--	0.215	PASS
Standard Curve (r)	0.99	--	--	0.999	PASS

KIT COMPONENTS :			
NAME	LOT NUMBER	EXP. DATE	VOLUME
CaCl ₂	N0254975	2018-02	(1x) 6 ml
CT SP Substrate	N0355360	2017-03	(1x) 12 ml
Factor Reagent	N0455731	2017-08	(4x) 3 ml
CT SP Buffer, pH 7.3	N0355255	2017-03	(1x) 20 ml
CT SP Phospholipid	N0355634	2017-03	(1x) 2 ml

PREPARED BY:

NPATEL

DATE:

05/20/15

APPROVAL SIGNATURE:

D7EMCBA

DATE:

05/20/15