

CERTIFICATE OF ANALYSIS





DiaPharma Group, Inc.
8948 Beckett Road, West Chester, OH 45069

Product:	DiaPharma Factor X kit
Item Number:	DPGFX
Vendors:	R ² Diagnostics, Chromogenix
Lot #:	DPGFX040
Exp. Date:	12/2015

Kit Components:			
Name	Lot Number	Exp. Date	Volume (1 vial each)
FXa Chromogenic Substrate	N1036824	10/2016	25 mg
Russell's Viper Venom	DR-903	12/2015	<1 mg
CaCl ₂	DC-903	12/2015	20 ml
Tris Buffer	DB-903	12/2015	100 ml

Test Description	Specifications		Results		Disposition
	Min	Max	Units	Value	
Slope	1.0e-3	4.0e-3	NA	3.079e-03	Pass
Intercept	-0.09	0.09	NA	5.749e-03	Pass
Regression Coefficient	R > 0.990		NA	0.999	Pass

Prepared By:  Date: 06/23/2014

Approved By:  Date: 6/23/2014