

CERTIFICATE OF ANALYSIS





DiaPharma Group, Inc.
8948 Beckett Road, West Chester, OH 45069

Product:	DiaPharma Factor X kit
Item Number:	DPGFX
Vendors:	R ² Diagnostics, Chromogenix
Lot #:	DPGFX042
Exp. Date:	03/2017

Kit Components:			
Name	Lot Number	Exp. Date	Volume (1 vial each)
FXa Chromogenic Substrate	N1143454	11/2017	25 mg
Russell's Viper Venom	DR-931	03/2017	<1 mg
CaCl ₂	DC-931	03/2017	20 ml
Tris Buffer	DB-931	03/2017	100 ml

Test Description	Specifications		Results		Disposition
	Min	Max	Units	Value	
Slope	1.0e-3	4.0e-3	NA	2.963e-03	Pass
Intercept	-.09	.09	NA	8.689e-03	Pass
Regression Coefficient	R > 0.990		NA	0.998	Pass

Prepared By:  Date: 10/08/2015

Approved By:  Date: 10/26/15