

**CERTIFICATE OF ANALYSIS**



DiaPharma Group, Inc.  
8948 Beckett Road, West Chester, OH 45069

<b>Product:</b>	DiaPharma Factor X kit
<b>Item Number:</b>	DPGFX
<b>Vendors:</b>	R <sup>2</sup> Diagnostics, Chromogenix
<b>Lot #:</b>	DPGFX039
<b>Exp. Date:</b>	06/2015

<b>Kit Components:</b>			
<b>Name</b>	<b>Lot Number</b>	<b>Exp. Date</b>	<b>Volume (1 vial each)</b>
FXa Chromogenic Substrate	N0333612	03/2016	25 mg
Russell's Viper Venom	DR-893P	06/2015	<1 mg
CaCl <sub>2</sub>	DC-893P	06/2015	20 ml
Tris Buffer	DB-893P	06/2015	100 ml

<b>Test Description</b>	<b>Specifications</b>		<b>Results</b>		<b>Disposition</b>
	Min	Max	Units	Value	
Slope	1.0e-3	4.0e-3	NA	3.115e-03	Pass
Intercept	-.09	.09	NA	3.518e-03	Pass
Regression Coefficient	R > 0.990		NA	0.999	Pass

Prepared By:  Date: 12/11/2013

Approved By:  Date: 12/16/2013