

CERTIFICATE OF ANALYSIS

**CHROMOGENIX**

PRODUCT:	S-2222, 25 mg Substrate	PRODUCT #:	82031639
EXP. DATE:	2016-07	LOT #:	N0333613
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	820316

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	25.7	Pass
FREE p-NITROANILINE	0	0.5	mol %	0.1	Pass

DESCRIPTION	LOT NUMBER	EXP. DATE	VOLUME
S-2222	N0333613	2016-07	6 ml

PREPARED BY: M. Oettinger M. Oettinger      DATE: 4/22/2013 4.22.13

APPROVAL SIGNATURE: D. Tencza P. Ronza      DATE: 4/22/2013 4/22/13

Form QCM 820316-2 Rev. 01/09