

CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT:	S-2222, 25 mg Substrate	PRODUCT #:	82031639
EXP. DATE:	2017-09	LOT #:	N0540429
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	820316

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	24.8	PASS
FREE p-NITROANILINE	0	0.5	mol %	0	PASS

DESCRIPTION	LOT NUMBER	EXP. DATE	VOLUME
S-2222	N0540429	2017-09	6 ml

PREPARED BY: M. Oettinger Maria Oettinger DATE: 6/17/2014 6-17-14

APPROVAL SIGNATURE: D. Tencza D. Tencza DATE: 6/17/2014 6/17/14

Form QCM 820316-2 Rev. 01/09