

CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT: S-2238, 25 mg Substrate	PRODUCT #: 82032439
EXP. DATE: 2016-03	LOT #: N0333448
VENDOR: INSTRUMENTATION LABORATORY	ASSAYED PER QCM: 820324

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	25.5	Pass
FREE p-NITROANILINE	0	0.5	mol %	0.1	Pass

DESCRIPTION	LOT NUMBER	EXP. DATE	VOLUME
S-2238	N0333448	2016-03	6 ml

PREPARED BY: M. Oettinger *M. Oettinger* DATE: 3/29/2013 *3-29-13*

APPROVAL SIGNATURE: D. Tenca *D. Tenca* DATE: 3/29/2013 *4/1/13*

Form QCM820324-2 Rev. 01/09