

CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT: S-2251, 25 mg Substrate	PRODUCT #: 82033239
EXP. DATE: 2015-08	LOT #: N0821228
VENDOR: INSTRUMENTATION LABORATORY	ASSAYED PER QCM: 820332

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	25.3	PASS
FREE p-NITROANILINE	0	0.5	mol %	0	PASS

DESCRIPTION	LOT NUMBER	EXP. DATE	VOLUME
S-2251	N0821228	2015-08	3 ml

PREPARED BY: M. Oettinger *M. Oettinger* DATE: 9/10/2012 *9-10-12*

APPROVAL SIGNATURE: D. Tencza *D. Tencza* DATE: 9/10/2012 *9/12/12*