

CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT: S-2302, 25 mg Substrate	PRODUCT # : 00082034039
EXP. DATE: 2017-05	LOT # : N0148448
VENDOR: INSTRUMENTATION LABORATORY	ASSAYED PER QCM: 820340

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	24.5	PASS
FREE p-NITROANILINE	0	0.5	mol %	0.0	PASS

DESCRIPTION	LOT NUMBER	EXP. DATE	VOLUME
S-2302	N0148448	2017-05	3 mL

PREPARED BY: *M. Oettinger* DATE: 02/06/2014

APPROVAL SIGNATURE: *D. Tencza* DATE: 02/06/2014