CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT:	S-2302, 25 mg Substrate	PRODUCT # :	00082034039
EXP. DATE:	2017-09	LOT # :	N0540427
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	820340

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	DISPOSITION
	MIN	MAX	UNITS		PASS/FAIL
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	24.0	Pass
FREE p-NITROANILINE	0	0.5	mol %	0.0	Pass

DESCRIPTION	LOT NUMBER	EXP.DATE	VOLUME
S-2302	N0540427	2017-09	3 mL

PREPARED BY: *M. Oettinger* DATE: 6-Jun-2014

APPROVAL SIGNATURE: *D. Tenezza* DATE: 6-Jun-2014

Form QCM00820340-2 Rev. 03/09

526 Route 303, Orangeburg, New York 10962, Telephone 845-365-8000